

2020-02-28

SikaMox 20/40
 Revison: 1.1
 S/N SikaMox 20/40: 20-02-042

Calibration example

Object: EDM turbine Flowmeter S071
 S/N Meter: 0420154

Set(1-6)	Units	1	2	3	4	5	6
Reading KV	Kg	99,90	100,20	112,00	100,15	101,15	
Reading PM	Litres	100,18	100,49	112,34	100,34	101,47	
PM converted to mass	Kg	100,00	100,31	112,14	100,16	101,29	
Flow rate	l/min	20	25	30	35	40	
Measurement error	%	0,1	0,11	0,12	0,01	0,14	

Unit: Litres Pulses Kg

Test liquid: Water Exxol D80

Water temperature: 20 °C

Density: 0,998203 kg/dm³

Traceability: KK1500 = Calibration Vessel 1 500 l Evidence no. MTq F012636-K02 Abbreviation
 KK200 = Calibration Vessel 200 l Evidence no. MTq F012636-K03 PM = Test Object
 KK25 = Calibration Vessel 25 l Evidence no. MTq F012636-K01 MM = Master Meter
 KV = Calibrated Scale S/N B823952382 / 91907 KK = Calibration Vessel
 MM1 = Master Meter DN 50 S/N 77292 CL KV = Calibrated Scale
 MM2 = Master Meter DN 50 S/N 189413
 MM3 = Master Meter DN 80 S/N 9014
 MM4 = Master Meter DN 100 S/N 30167
 MM5 = Master Meter DN 100 S/N 138539
 Other =

Master Meter (MM) is calibrated against the calibration vessel (KK) with evidence no. MTV F613218-02 for 1500 liter vessel, certificate no. MTV F613218-03 for 200 liter vessel and evidence no. MTV F613218-01 for 25 liter vessel. Calibrated scale (KV) is verified according to EN 45501:2015 with EC type approval and test certificate T11213.

uncertainty: The margin of error of calibration equipment and calibration method is less than 0.25%.

Remark:

Pär Broberg
 Test Manager



Patrik Fryklín
 Quality Manager

